



## Declaration of Conformity

Manufacturer: COBI Rehab ApS  
Fuglebækvej 1 D  
DK-2770 Kastrup, Danmark  
Tel. +45 7025 2522  
E-mail: [cobi@cobi.dk](mailto:cobi@cobi.dk)

The undersigned declares that the product named and listed on this certificate is constructed and manufactured in accordance with the essential requirements of The Regulation (EU) 2017/745 of the European Parliament and of the Council of 5. April 2017 relating to Medical Devices Class 1.

Risk analysis has been executed in accordance with EN ISO 14971:2012 Medical devices  
- Application of risk management to medical devices.

The devices comply to the following standards:

EN 12182 Assistive products for persons with disability. General requirements and test methods  
ISO 10535  
EN 55011:2016, EN 61000-3-2:2014, EN 61000-3-3:2013,  
EN 61000-4-2:2009, EN 61000-4-3:2006+A1+A2, EN 61000-4-4:2012, EN 61000-4-5:2014, EN 61000-4-6:2014,  
EN 61000-4-8:2010, EN 61000-4-11:2004.  
EN 60601-1:2006+A1  
EN 50581:2012  
EN 60601-1-2: 2015, Edition 4.0  
EN 60601-2-52:2010

Product name:	XXL-Rehab Shower Commode Rise and Tilt
Item No.:	0157-066-000
Risk Class:	I
Intended use:	Assistive devices for persons with special needs
Basic UDI-DI:	5740000100005N4
SRN:	DK-MF-000001370

Kastrup, 17-02-2021

Claus Nymark, Director